

CALIBRATION CERTIFICATE

ACCORDING TO Q/320684FHN13-2007

Commodity: LOOP 1UL AND LOOP 10UL

MATERIAL:ABS

THE LOOP HEAD SIZE accord with standard GB/T14486

IUL HEAD DIAMETER 2MM WHICH CAN TAKE IUL LIQUID

10UL HEAD DIAMTER 5.8MM WHICH CAN TAKE 10UL LIQUID

LOOP HEAD :BAR NOT ALLOWED.

2 BULB: BIG BULB NOT ALLOWED

3 COLOR: NO COLOR distinctness

STERILE BY GAMMA RAY ACCORDING TO GB 15980

ACCORDING TO THE STANDARD:

THE GOODS PASS THE TEST INVOICE NO: RN2024-149

LOT NO: Lot R202408

QTY: 1UL: 3000PACKS

10UL: 3000PACKS

Date 2024-08-25

浙江理能医疗器械有限公司
ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD

