

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Ceftazidime*-avibactam CZA 14 µg	081924140	2027.08.19

Ref.

9206 – 9206/1 – 9206/2

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	24-30	28
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	21–27	24
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18–24	21

Batch Release

Approved

Date

26.08.2024

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.